

## Nitrile exam gloves with excellent dexterity along with enhanced comfort, fit and feel

- **Elevated comfort:** These thin, black nitrile medical exam gloves offer enhanced dexterity and tactility in a thin, yet strong design
- **Black nitrile:** Distinctive black color provides contrast while also hiding the appearance of oils, bodily fluids and other stains
- **ERGOFORM™ Technology:** ERGOFORM™ Ergonomic Design Technology reduces muscle strain during repetitive tasks, improving worker comfort and productivity
- **Optimized grip:** Textured on the fingertips for secure handling of instruments and materials



 **ERGOFORM™**  
Ergonomic Design Technology

### Industries

- Food Service
- Cosmetics
- Tattoo
- Automotive Aftermarket
- Chemical
- Emergency Medical Services
- Healthcare
- Law Enforcement
- Life Sciences
- Dental
- Machinery and Equipment
- Metal fabrication

### Recommended For

- General dentistry
- Standard patient examinations
- Emergency services
- Catering/Meal preparation
- General Purpose Auto Aftermarket
- Equipment repair and maintenance
- Laboratory analysis
- Cosmetic Procedures
- Tattoo

TECHNICAL DATA SHEET

	Product Information
Material	Nitrile
External Glove Surface	Textured Fingers
Audit Standards	21 CFR 820, ISO 13485, ISO 9001
Product Certification	Medical Device Regulation (EU) 2017/745 Class I, Personal Protective Equipment Regulation (EU) 2016/425 Category III risks, EU Food Contact Approved, EC 2023/2006, Anvisa 81496190021 / CA 48.379, INVIMA 2015DM-0013417, CE 0493, EN 455 1-4, EN ISO 21420:2020, EN ISO 374-1:2016 Type B KPT, EN ISO 374-5:2016 Virus, EN 16523-1, ISO 18889:2019, ISO 9001, ISO 13485, UKCA, US QSR/FDA 510(k) Medical Examination Grade, FDA21 CFR 177-2600-US Food Contact Approved, ASTM D6319, EN 421:2010
Packaging Overview	Case=1000 gloves/10 dispensers; Dispenser=100 gloves 2XL: Case=900 gloves/10 dispensers; Dispenser=90 gloves
Storage Instructions	Keep out of direct sunlight; store in a cool and dry place. Keep away from sources of ozone or ignition.
Country Of Origin	China, Malaysia
Product Segmentation	High Touch
Antistatic	No
Silicone Free	No

PHYSICAL PROPERTIES

	Typical Values		Testing Method
Length (mm/inches)	245 / 9.6		ASTM D6319,EN 455-2
Freedom from Holes (Inspection level I)	1.5 AQL		ASTM D5151,EN 455-1
Palm Thickness (mm/mils)	0.07 / 2.8		ASTM D6319,EN 455-2
Finger Thickness (mm/mils)	0.11 / 4.3		ASTM D6319,EN 455-2
	BEFORE AGING		AFTER AGING
Ultimate Tensile Strength (MPa)	≥ 18	≥ 16	ASTM D412 & D573
Elongation at Break (%)	≥ 500	≥ 400	ASTM D412 & D573
Force at break (N)	≥6	≥6	EN 455

ORDERING INFORMATION

Size	XS (5.5 - 6)	S (6.5 - 7)	M (7.5 - 8)	L (8.5 - 9)	XL (9.5 - 10)	XXL (10.5 - 11)
Product Code	93732060	93732070	93732080	93732090	93732100	93732110

Performance Standards and Regulatory Compliance



For additional information visit us at [www.ansell.com](http://www.ansell.com), or call us at

**Europe, Middle East & Africa Region**  
Ansell Healthcare Europe NV  
T: +32 (0) 2 528 74 00  
F: +32 (0) 2 528 74 01  
  
**Asia Pacific Region**  
MY +603 8310 6688  
apac.medical@ansell.com  
CN: +86 21 38275000  
infochina@ansell.com  
JP: +813 5549 8151  
info.medical.jp@ansell.com

**North America Region**  
Ansell Healthcare Products LLC  
T: +1 800 800 0444  
F: +1 800 800 0445  
  
**Latin America & Caribbean Region**  
BR +55 (11) 3356-3100  
latam.medical@ansell.com  
MX: +52 442 296 2050  
latam.medical@ansell.com  
CO: +57 1 288 3247  
latam.medical@ansell.com

**Australia**  
Ansell Limited  
T: +61 1800 337 041  
F: +61 1800 803 578  
  
**Russia**  
Ansell PYC  
Ten. +7 495 258 13 16

Technology



Ansell, ® and ™ are trademarks owned by Ansell Limited or one of its affiliates. US Patented and US and non-US Patents Pending: [www.ansell.com/patentmarking](http://www.ansell.com/patentmarking) © 2025 Ansell Limited. All Rights Reserved.

Neither this document nor any other statement made herein by or on behalf of Ansell should be construed as a warranty of merchantability or that any Ansell product is fit for a particular purpose. Ansell assumes no responsibility for the suitability or adequacy of an end user's selection of gloves for a specific application.

