

# SafeGrip<sup>™</sup> SG-375 Thick, Durable Latex Exam Glove with Extended Cuff

# Extra-thick latex medical exam gloves, offering durable protection for heavy-duty applications

- Extra defense: MICROFLEX<sup>®</sup> SafeGrip<sup>™</sup> SG-375 latex medical exam gloves are 3x thicker than standard disposable latex gloves, ensuring greater resistance against snags and tears
- **Comfortable fit:** These thick latex gloves' soft, elastic formulation delivers comfort and flexibility
- Added protection: With an extended cuff, these gloves offer workers enhanced forearm and wrist protection
- Enhanced grip: These exam gloves' textured fingertips enhance grip performance
- Easy donning: Donning is made easy thanks to the gloves' double-chlorinated design
- Certified performance: They satisfy NFPA 1999 emergency medical glove requirements and are tested against chemotherapy drugs

CAUTION: This product contains natural rubber latex, which may cause allergic reactions

## Industries

- Automotive
- Automotive Aftermarket
- Life Sciences
- Healthcare

- Recommended For
- Inspection, selecting, checking parts
- Assembly and inspection of components
- Picking and fastening components
- Equipment repair and maintenance
- General Purpose Auto Aftermarket
- Body panel repair and replacement
- Tire rotation & changes
- Animal care feeding and cleanup
- Maintaining lab furniture and equipment
- Sample taking and processing
- Raw material sample collection
- Extra protection over the wrist and arm
- Higher-risk clinical applications
- Tested for use with chemotherapy drugs





## **TECHNICAL DATA SHEET**

	Product Information
External Glove Surface	Textured Fingers
Audit Standards	EN ISO 13485:2012
Product Certification	Medical Device Regulation (EU) 2017/745 Class I, ASTM D3578, ASTM D6978, EU Food Contact Approved, EN ISO 21420:2020, FDA21 CFR 177-2600-US Food Contact Approved, NFPA 1999:2018, US QSR/FDA 510(k) Medical Examination Grade
Packaging Overview	50 gloves per dispenser 10 dispensers per case 500 gloves per case
Storage Instructions	Keep out of direct sunlight; store in a cool and dry place. Keep away from sources of ozone or ignition.
Country Of Origin	Malaysia
Product Segmentation	High Risk
Antistatic	Not Tested
Silicone Free	No

## **PHYSICAL PROPERTIES**

	Typical Values		Testing Method
Length (mm/inches)	295 /	295 / 11.6	
Freedom from Holes (Inspection level I)	1.5 AQL		ASTM D5151,EN 455-1
Palm Thickness (mm/mils)	0.28 / 11		ASTM D3767,EN 420
Finger Thickness (mm/mils)	0.36 / 14.2		ASTM D3767,EN 420
	BEFORE AGING	AFTER AGING	
Ultimate Tensile Strength (MPa)	≥ 34	≥ 21	ASTM D412 & D573
Elongation at Break (%)	≥ 850	≥ 750	ASTM D412
Force at break (N)	≥ 15	≥ 10	EN 455-2

## **ORDERING INFORMATION**

Size	S (6.5 - 7)	M (7.5 - 8)	L (8.5 - 9)	XL (9.5 - 10)
Product Code	SG-375-S	SG-375-M	SG-375-L	SG-375-XL

#### **Performance Standards and Regulatory Compliance**



For additional information visit us at www.ansell.com, or call us at

#### Europe, Middle East & Africa Region

Ansell Healthcare Europe NV T: +32 (0) 2 528 74 00 F: +32 (0) 2 528 74 01

#### Asia Pacific Region MY +603 8310 6688 apac.medical@ansell.com CN: +86 21 38275000

infochina@ansell.com JP: +813 5549 8151 info.medical.jp@ansell.com North America Region Ansell Healthcare Products LLC T: +1 800 800 0444 F: +1 800 800 0445

#### Latin America & Caribbean Region BR +55 (11) 3356-3100 latarn.medical@ansell.com MX: +52 442 296 2050 latarn.medical@ansell.com CO: +57 1288 3247 latarn.medical@ansell.com

Australia Ansell Limited T: +61 1800 337 041 F: +61 1800 803 578

**Russia** Ansell PYC Ten. +7 495 258 13 16

Ansell, <sup>®</sup> and <sup>™</sup> are trademarks owned by Ansell Limited or one of its affiliates. US Patented and US and non-US Patents Pending: www.ansell.com/patentmarking © 2025 Ansell Limited. All Rights Reserved.

Neither this document nor any other statement made herein by or on behalf of Ansell should be construed as a warranty of merchantability or that any Ansell product is fit for a particular purpose. Ansell assumes no responsibility for the suitability or adequacy of an end user's selection of gloves for a specific application.

